


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-1M2-P

S.N.: 9057

Date: SEPT 20 1999



- a) Output signal amplitude:  
 $\geq +60 \text{ Volts (50}\Omega)$
- b) Pulse width:  
 $\leq 2 \text{ ns}$
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
 $+135 \pm 5 \text{ VDC}$   
3 mA MAX.

50  $\Omega$  ATTEN  
12 32 V/DIV  
PRF = 300 kHz