


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVIR-1-C-PN-CA*

S.N.: *9079*

Date: *NOV 22 1999*



- a) Output signal amplitude:
0 TO ±200V (R_L = 50Ω)
- b) Pulse width:
1 TO 5NS
- c) Rise time:
≤ 0.5NS
- d) Fall time:
≤ 1.0NS
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz

60 db ATTN

100 VOLTS/DIV

2 NS/DIV

PRF = 50 KHz