


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AUP-AN-14V2-C-0A-P*

S.N.: *9081*

Date: *DEC 21 1999*



- a) Output signal amplitude:
0 to +20V (70.50m)
- b) Pulse width:
0.3 to 2.0ns
- c) Rise time:
≤ 100ps
- d) Fall time:
≤ 200ps
- e) PRF:
0 to 1.0 MHz
- f) Jitter, stability:
- g) Prime power:
OK

40 dB ATTEN

2 10 VOLTS/DIV

TOP PW NORM MIN

NOI PW NORM MAX

PRF = 100 kHz

120/240V
50-60 MHz
~~*100*~~