

PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV-110B-PS-LTA*

S.N.: *9105*

Date: *NOV 23 1999*


- a) Output signal amplitude:
0 TO ±200VOUTS (RL > 50K)
- b) GAIN:
X1 TO X100
- c) Rise time:
≤ 145 (V₀₋₁)
- d) Fall time:
≤ 145
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power: *120/240V*
50-60Hz
- h) OFFSET: *0 TO ±150V*



9105
 $R_L = 50K$

V_{out}

V_{in}



9105
 $R_L = 50k$

V_{out}

V_{in}

