


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-2-C-PN-05*  
 S.N.: *9153*  
 Date: *FEB 22 2000*

- a) Output signal amplitude: *0 TO ±10 Amps*
- b) Pulse width: *TO 5 μs*  
*0 TO 15 NS*
- c) Rise time: *≤ 1 NS*
- d) Fall time: *≤ 2 NS*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*



*200 P out*  
*50 I N out* }  $\approx 32 \text{ V/DIV.}$   
*PRF = 20 KHz.*

*50-60 V/DIV.*  
*[Signature]*

