


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MVN-1-C-P*
S.N.: *9164*
Date: *FEB 2 2000*



- a) Output signal amplitude:
0 TO + 5 Volts (TO 50mV)
- b) Pulse width:
0.2 TO 1.0 NS
- c) Rise time:
≤ 100ps
- d) Fall time:
≤ 100ps
- e) PRF: *50 TO 250 MHz.*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60 Hz.

40 db ATTEN
5 VOLTS/DIV.
TRP 2 NS/DIV
PROT. AS TOP BUT 200ps/DIV.