


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMM-2-C-N*

S.N.: *9168*

Date: *MAR 6 2000*



- a) Output signal amplitude: *0 TO -15 V (To 50m)*
- ① b) Pulse width: *0.2 TO 2.0 NS*
- ② c) Rise time: *≤ 100 ps*
- ③ d) Fall time: *≤ 135 ps*
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60 MHz

NY

40 db ATTEN.: 10 U/DIV
 ① *5 NS/DIV, PRF NORM MAX.*
 ② *AS ① BUT 1 NS/DIV*
 ③ *1 NS/DIV, PRF ≈ 2.5 MHz*