


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-M-1-CP-05-M*

S.N.: *9171*

Date: *DEC 22 1999*



- a) Output signal amplitude:
0 TO +10V_{0.43} (50%)
- b) Pulse width:
0.2 TO 4.0ns
- c) Rise time:
≤ 100ps
- d) Fall time:
≤ 135ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V

40 dB ATTEN

10 V/VTB 1DIV

PW ≈ MAX

AND ≈ MIN

PRF = 100 kHz.

50-60 Hz