


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-MV-15-C-N-05*

S.N.: *9190*

Date: *MAR 13 2000*



- a) Output signal amplitude:
0 TO -5V (50Ω)
- b) Pulse width:
0.2 TO 4.0 ns
- c) Rise time:
≤ 60 ps (20-80%)
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz

40 dB ATTEN
5 V/DIV
TOP: 1 ns/DIV
BOT: 100 ps/DIV
(CLEARING EXHIBIT)