

PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-151A-C-LT1


S.N.: 9209

Date: APRIL 2 2000

- a) Output signal amplitude:
0 to $\pm 200V$ ($R_L \geq 50K$)
- b) Pulse width:
SINE, TRIANGLE, SQUARE
AND DC OFFSET
- c) Rise time:
 $\leq 1\mu s$ (V_{0-p})
- d) Fall time:
 $\leq 1\mu s$
- e) PRF:
0 to 100 KHz.
- f) Jitter, stability:
OK
- g) Prime power: 120/240V
50-60 Hz.
- h) DC OFFSET:
* 0 to $\pm 150V$.



9209
 $R_c = 50K$



9209
 $R_L = 50K$

