


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-2C-PEPIA*
 S.N.: *9251*
 Date: *MAY 2 2000*



- a) Output signal amplitude: \checkmark
0 TO +10 V (TO 50 Ω)
- b) Pulse width: \checkmark
5 TO 100 NS
- c) Rise time: \checkmark
 ≤ 100 ps
- d) Fall time: \checkmark
 ≤ 135 ps
- e) PRF: \checkmark
0 TO 100 MHz
- f) Jitter, stability: \checkmark
OK
- g) Prime power: \checkmark
120/240 V
50-60

40 dB ATTEN
-- 10 V/DIV
TOP 5 NS/DIV) BOTH.
MID 200ps/DIV (RISE TIME)
BOT 200ps/DIV (FALL TIME)
PRF \approx 100 kHz

[Signature]