

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AN-110B-PS*
S.N.: *9257*
Date: *MAR 6 2000*

- a) Output signal amplitude:
± 200 V MAX
- b) *GAIN*: *R_L ≥ 50K:* *X1 TO X100*
- c) Rise time: *≤ 145*
- d) Fall time: *(V_{OP})*
≤ 145
- e) PRF: *DC TO 100KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*
50-60Hz



9257

$R_L = 50 \text{ K.}$

