


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP.-2-B-PN*

S.N.: *9305*

Date: *MAY 17 2000*



- a) Output signal amplitude:
0 TO ± 10 VOLTS (± 50%)
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/200V
50-60 Hz

40 dB ATTEN. ∴ 100.0 NS/DIV

- ① *Pout 5 ns/div*
- ② *Pent 500ps/div (RISE TIME)*
- ③ *Nout 5 ns/div*
- ④ *Nent 500ps/div (RISE TIME)*

[Signature]