

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-C-TRWA*

S.N.: *9306*

Date: *MAY 19 2000*


- a) Output signal amplitude:
0 TO +14 VOLTS TO
- b) Pulse width: *$R_L \geq 7 \Omega$*
10 μ S TO 10 MS
- c) Rise time:
50% MAX DUTY CYCLE
 ≤ 100 NS
- d) Fall time:
 ≤ 100 NS
- e) PRF: *0 TO 1 KHz*
50% MAX DUTY CYCLE
- f) Jitter, stability:
OK
- g) Prime power: *120/240V*
50-60Hz
~~MB~~

9306

$$R_L = 7\Omega$$

$$V_{out} = 14 \text{ V RMS}$$

$$PRF = 300 \text{ Hz}$$



9306
 $R_c = 7\Omega$
RISE TIME W/INSTRUM.

