


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-3 P*
S.N.: *9345*
Date: *MAY 26 2000*



- a) Output signal amplitude:
0 to +20V (to 50 Ω)
- b) Pulse width:
8 to 100 ns
- c) Rise time:
 ≤ 200 ps
- d) Fall time:
 ≤ 300 ps
- e) PRF:
0 to 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24V DC

250 mA
(max)

[Signature]

40 db ATTEN
10 V/CM
TOP 5 ns/CM
BOT 500 ps/CM
(RISE TIME)

PRF = 100 kHz