


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-3-P*
S.N.: *9346*
Date: *MAY 26 2000*



- a) Output signal amplitude:
0 TO +20 VOLT
- b) Pulse width: *(50 ns)*
8 TO 100 ns
- c) Rise time:
≤ 200 ps
- d) Fall time:
≤ 300 ps
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
+24 VDC
250 mA (MAX)

460 dB ATTEN
10 VOLT/DIV
5 ns/DIV
500 ps/DIV (RISE TIME)
PRF = 100 kHz