


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVIR-2-CPN-05-PWT-AT


S.N.: 9363

Date: JUNE 20 2000



- a) Output signal amplitude: 0 TO ±200 VOLTS
- b) Pulse width: (TO 50N)
2 TO 10 N
- c) Rise time: ≤ 1 N
- d) Fall time: pw (min) ≤ 1 N
pw (max) ≤ 2 N
- e) PRF: 0 TO 50 KHz

ⓐ 60 db ATTEN.: 100V/DIV
PW NORMAL MODE.



- f) Jitter, stability: OK
- g) Prime power: 120/240V
50-60Hz

[Handwritten signature]

ⓑ 60 db ATTEN.: 100V/DIV
PW NORMAL MODE
PRF = 20 KHz