


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-15.1B-C*
S.N.: *9364*
Date: *MAY 15 2000*


- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) Pulse width: *(R_L ≥ 50K)*
SINE, SQUARE, TRIANGLE
- c) Rise time: *+DC OFFSET*
≤ 1.0 μs (V₀₋₉)
- d) Fall time:
≤ 1.0 μs
- e) PRF:
DC TO 100 KHz
- f) Jitter, stability:
OK.
- g) Prime power: *120/240V*
50-60 Hz.

[Signature]

9364



9364



9364.

