


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AUP-AV-4N3-C-N*

S.N.: *9383*

Date: *JUN 13 2000*



- a) Output signal amplitude: *0 TO -40V (TO 50N)*
- b) Pulse width: *0.4 TO 4.0NS*
- c) Rise time: *≤ 150 ps*
- d) Fall time: *≤ 250 ps*
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

① output TO 40 db.

② *AA-51-KSA MFI OUT*
TO 20 db.

PRF = 100 kHz

120/240V
50-60Hz