

PULSE GENERATOR
PERFORMANCE CHECK

Model:

AV-110B-PS

S.N.:

9402

Date:

JUNE 7 2000

- a) Output signal amplitude:
 $\pm 200V$ MAX
- b) GAIN: X 1 TO X 100
- c) Rise time: $\leq 1\mu s$ (V_{DP})
- d) Fall time:
 $\leq 1\mu s$
- e) PRF: DC TO 100KHz
- f) Jitter, stability: 6K
- g) Prime power:

170/240V

50-60Hz



9402
 $R_L = 50K$

OUT

IN

