


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PUHF-1*  
S.N.: *9412*  
Date: *DEC 21 2000*



- a) Output signal amplitude: *≥ -10V*
- b) Pulse width: *≤ 130ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *100 ± 5 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*0.5 WATT (15 Vpp)*  
*100 MHz CW*

*40 db ATTEN*  
*10 VOLT/DIV*  
*TRF: 2ns/DIV*  
*BOT: 200ps/DIV*