


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-9A-C-N-MD*

S.N.: *9437*

Date: *AUG 9 2000*



- a) Output signal amplitude:
0 TO - 200 mA
- b) Pulse width:
0.4 TO 4.0 NS
- c) Rise time:
≤ 200 ps
- d) Fall time:
≤ 200 ps
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power:

TOP MAINFRAME OUT
40 db (NO AX-SI)
10V/DIV, 1NS/DIV
PRF = 100 KHz.

BOT AS TOP BUT AX-SI
INSTALLED AND ME OUT
TO 20 db. IN 459A DIODE
USED AS A SIMULATED LOAD.

120/240 V

50-60 Hz

[Signature]