


PULSE GENERATOR
PERFORMANCE CHECK

Model: *NR-1-PW-C-P-0A*


S.N.: *9485*

Date: *SEPT 6 2000*

- a) Output signal amplitude:
0 TO + 200 VOLTS
- b) Pulse width: *($t_r \geq 50\text{ns}$)*
50NS TO 5.0uS
- c) Rise time:
 $\leq 10\text{NS}$
- d) Fall time: *(20-80%)*
 $\leq 10\text{NS}$
- e) PRF: *0 TO 1 K Hz.*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60 Hz.



9485
NARROW PULSE
 $R_c = 50\Omega$



Q485
WIDE PULSE
 $R_L = 50\Omega$

