


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-EBBIE*

S.N.: *9495*

Date: *NOV 9 - 2008*



- a) Output signal amplitude: ✓
≥ +10 V₀ (70583)
- b) Pulse width: *1.5 NS FWHM* ✓
- c) Rise time: *≤ 100 ps* ✓
- d) Fall time: *≈ 3 NS* ✓
- e) PRF: *0 TO 2 MHz* ✓
- f) Jitter, stability: *OK* ✓

- g) Prime power:
 - a) *+15 VDC* ✓
160 mA
 - b) *-5 VDC*
10 mA

40 db ATTEN
10 V₀ / DIV

TOP: 1 NS / DIV

BOT: 100 ps / DIV

PRF = 2.0 MHz