


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-MPT-EGG1E*

S.N.: *9510*

Date: *Dec. 14 - 2008*



- a) Output signal amplitude:  
 *$\geq +10 \text{ V}_{\text{OVP}}$  (70583)*
- b) Pulse width:  
*1.5 NS FWHM*
- c) Rise time:  
 *$\leq 100 \text{ ps}$*
- d) Fall time:  
 *$\approx 3 \text{ NS}$*
- e) PRF:  
*0 TO 2 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:
  - a) *+15 VDC*  
*160 mA*
  - b) *-5 VDC*  
*10 mA*

*460 db ATTEN*  
*10 V OVP / DIV*  
*TOP: 1 NS / DIV*  
*BOT: 100 ps / DIV*

*PRF = 2.0 MHz*