

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-P-D-M*

S.N.: *9559*

Date: *OCT 18 2000*



- a) Output signal amplitude:
0 TO +15V (TO 50m)
- b) Pulse width:
0.2 TO 2.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*+24 VDC,
500 mA (MAX)*

*40 db ATTEN
10 V/DIV
PRF = 10 MHz*

A