


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-2-N-D-M*

S.N.: *9560*

Date: *OCT 18 2000*



- a) Output signal amplitude:
0 to -15V (to 50Ω)
- b) Pulse width:
0.2 to 2.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF: *0 to 25 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
124 VDC
500 mA (max)

40 db ATTN
∴ 10V/DIV
PRF ≈ 10 MHz