

PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-5B-PS-PN-0A-0W
S.N.: 9619
Date: JAN 22 2001

- a) Output signal amplitude:
0 to $\pm 500V$ ($R_L \geq 50\Omega$)
- b) Pulse width:
0.1 to 100 μs
(1% max duty cycle)
- c) Rise time:
 $\leq 30N$
- d) Fall time:
 $\leq 30N$
- e) PRF: 0 to 10 kHz
(1% max duty cycle)
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz.





(A)

9619

P out.

$R_c = 50\Omega$

NARROW PULSE.



ⓑ

9619
N OUT
 $R_L = 50\Omega$
WIDE PULSE

