


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-1-C-P*

S.N.: *9655*

Date: *FEB 20 2001*



- a) Output signal amplitude:
0 TO +5 V RMS
- b) Pulse width: *(TO 50ns)*
0.2 TO 6.0ns
- c) Rise time:
≤ 100ps
- d) Fall time:
≤ 13ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz

40 db ATTEN

5 V/DIV

PRF ≈ 12 MHz

① *pw ≈ MIN*

② *pw MID RANGE*

③ *pw ≈ MAX*

[Signature]