


PULSE GENERATOR
PERFORMANCE CHECK

Model: AN M14-4-C-KA PB1

S.N.: 9678

Date: MAY 7 2001



- a) Output signal amplitude: 0 TO +100 VOLTS
- b) Pulse width: 1 TO 3.5 NS
- c) Rise time: ≤ 0.5 NS
- d) Fall time: 1 TO 3 NS
- e) PRF: 0 TO 2 MHz (CW), 0 TO 10 MHz (BURST)
- f) Jitter, stability: OK

- g) Prime power: 120/240V
50-60Hz.

60 dB ATTEN.
100 VOLTS/1DIV
PRF = 2.0 MHz
(CW).

[Signature]