

PULSE GENERATOR

PERFORMANCE CHECK


Model: *AV-110A-PS*

S.N.: *9719*

Date: *MAR 20 2001*

- a) Output signal amplitude:  
*0 TO ±500 VOLTS*
- b) GAIN: *(R<sub>L</sub> ≥ 100K)*  
*X1 TO X250*
- c) Rise time:  
*≤ 35 μS*
- d) Fall time:  
*≤ 35 μS*
- e) PRF: *DC TO 3 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:



*170/240V*  
*50-60 Hz*



(A)

9719

$h_c = 100K$



(B)

9719

