


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-C-OT-P-PN*

S.N.: *9728*

Date: *MAY 8 2001*



- a) Output signal amplitude:  
*0 to ± 5 VOLT (0.50V)*
- b) Pulse width:  
*0.2 to 6.0 ns*
- c) Rise time:  
*≤ 100 ps*
- d) Fall time:  
*≤ 135 ps*
- e) PRF:  
*0 to 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: *170/240V*

*40 db ATTEN*  
*5 VOLT/DIV*  
*PRF ≈ 10 MHz*

*50-60 MHz*