

PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-110B-PS
S.N.: 9748
Date: MAR 2 2001

- a) Output signal amplitude:
0 TO $\pm 200V$ ($R_L \geq 50\Omega$)
- b) GAIN: $X1$ TO $X100$
- c) Rise time:
 $\leq 1\mu s$ (V_{OP})
- d) Fall time:
 $\leq 1\mu s$ (V_{OP})
- e) PRF: DC TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60Hz
A6

9748

$R_L = 50K.$

V_{comp}

V_{in}

