


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ARKL-1-PS-EN-M.*

S.N.: *9731*

Date: *JUNE 22 2001*



- a) Output signal amplitude:  
*~ 200 V to 50m*
- b) Pulse width:  
*5 to 100 ns.*
- c) Rise time:  
*≤ 2 ns*
- d) Fall time:  
*≤ 3 ns*
- e) PRF:  
*0 to 5 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*60 dB ATTEN*  
*100 VOLTS/DIV*  
*PRF = 2 kHz*

*100 ± 10 VOLTS*  
*50 to 60 Hz*