


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-C-N

S.N.: 9794

Date: JUNE 6 2001



- a) Output signal amplitude:
0 TO -350 V (POSITIVE)
- b) Pulse width:
5 TO 10 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

60 dB ATTEN
∴ ≈ 100 VOLTS/DIV

TDR 10 NS/DIV

VBOT 2 NS/DIV (RISE TIME)

120/240V
50-60 Hz
[Signature]