


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM4-1-CP*

S.N.: *9301*

Date: *APRIL 17 2001*



- a) Output signal amplitude: *0 to +10V*
- b) Pulse width: *≤ 200 ps (20-80%)*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 25 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50/60 Hz

40 db ATTEN
- 10 VOLT/DIV
PRF = 10 MHz

[Signature]