


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-C-P*

S.N.: *9882*

Date: *MAY 22 2001*



- a) Output signal amplitude:
0 TO +10 V (7050 Ω)
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
 ≤ 100 ps
- d) Fall time:
 ≤ 135 ps
- e) PRF:
0 TO 1.0 MHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz.

40 dB ATTEN
.10 VOLTS/DIV
TOP 1 NS/DIV
BOT. 100 ps/DIV (RISE TIME)
PRF \approx 100 KHz.

120/240V
50-60 Hz.
