

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PRH-3-CP-WA-W*

S.N.: *9931*

Date: *OCT 10 2001*

- a) Output signal amplitude:  
*0 TO +3000 V (TO R. 2/100)*
- b) Pulse width:  
*200NS TO 2.5US*
- c) Rise time:  
*≤ 100NS*
- d) Fall time:  
*≤ 100NS*
- e) PRF:  
*0 TO 1 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*  
*[Signature]*

9931

