


PULSE GENERATOR
PERFORMANCE CHECK

Model: MVP-MV-1-P

S.N.: 9953

Date: OCT 15 2001



40 dB ATTEN.: 10 V/DIV

TOP: 1 NS/DIV

BOI: 100 ps/DIV
(RISE TIME)

PRF = 100 KHz

- a) Output signal amplitude:
0 TO +10 VOLTS (TO 50V)
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
250 mA (MAX)