


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-2-C-P*

S.N.: *9977*

Date: *DEC 10 2001*



- a) Output signal amplitude:
0 to +15V (to 50Ω)
- b) Pulse width:
0.2 to 2.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 to 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATTEN
10V 0.4ns 10DIV

PRF = 10 MHz

100 ± 10V 0.4ns

50 - 60 MHz