


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-EGGIE*

S.N.: *9978*

Date: *OCT 16 2007*



- a) Output signal amplitude: ✓
≥ +10 V₀ (705.3)
- b) Pulse width:
1.5 ns FWHM
- c) Rise time:
≤ 100 ps
- d) Fall time:
≈ 3 ns
- e) PRF:
0 TO 2 MHz
- f) Jitter, stability:
OK
- g) Prime power:
 - a) *+15 VDC*
160 mA
 - b) *-5 VDC*
10 mA

40 db ATTEN
10 V₀ / DIV
TOP: 1 ns / DIV
BOT: 100 ps / DIV
PRF = 2.0 MHz

[Signature]