

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *NRH-3-CN*

S.N.: *9985*

Date: *NOV 9 2001*

- a) Output signal amplitude:  
*0 to 3000 Volts (R<sub>L</sub> ≥ 10Ω)*
- b) Pulse width:  
*200 ns to 2.5 μs*
- c) Rise time:
- d) Fall time: *≤ 100 ns (20-80%)*  
*≤ 100 ns (20-90%)*
- e) PRF:  
*0 to 1 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 Volts*  
*50-60 Hz*  
*[Signature]*

9985

$R_L \approx 10K$

$PBF \approx 1K\Omega$

