


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-TA1-P-EAS-M1-T1*

S.N.: *9997*

Date: *NOV 23 2001*



- a) Output signal amplitude:  
*0 TO +100V (TO 50Ω)*
- b) Pulse width:  
*≤ 0.6 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF: *0 TO 200 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*60 dB ATTEN*

*100V/DIV.*

*PRF = 200 KHz*

*+15 VDC.*

*250 mA (max)*

*[Signature]*