


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVO-1-C-P-WH*

S.N.: *10803*

Date: *DEC 17 2003*



*Test for $R_L = 2.2 \Omega$, $C_L = 10,000 \mu\text{F}$
(7115 current probe) 1 kHz.*



*Test for $R_L = 0.55 \Omega$, $C_L = 500 \mu\text{F}$
7115 current probe. 1 kHz.*

- a) Output signal amplitude:
*+25 TO +100 AMPS
(R_L 0.22 TO 2.2 Ω)*
- b) Pulse width:
5 TO 25 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 TO 1 KHz.*
- f) Jitter, stability:
OK
- g) Prime power:
*100 - 240V
50 - 60 Hz*