


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVN-1-C-P-PN*

S.N.: *10806*

Date: *DEC 04 2003*



- a) Output signal amplitude: *0 to ±5 V (to 50Ω)*
- b) Pulse width: *0.2 to 10 ns*
- c) Rise time: *≤ 100 ps*
- d) Fall time: *≤ 135 ps*
- e) PRF: *50 to 250 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/280V*

*40dB ATTN.: 5 V/DIV*

*TOP 5 ns/DIV*

*MID 1 ns/DIV*

*BOT 200 ps/DIV*