


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AUG-2A-N*  
S.N.: *11009*  
Date: *OCT 22, 2004*



- a) Output signal amplitude: *0 TO +120 V*
- b) Pulse width: *≤ 1 NS*
- c) Rise time: *≤ 0.4 NS*
- d) Fall time: *≤ 0.4 NS*
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*60 dB ATTN*

*100 VOLT/DIV*

*PRF = 10 KHz*

*+15 VDC*

*300 mA (mA)*